



NOTES:

1. DIMENSIONS GIVEN ARE FOR ENDEVCO 7264-2000 AND 7264C-2000 ACCELEROMETERS ONLY.
2. THE TOP, FRONT, LEFT AND REAR MOUNTS ARE INTERCHANGEABLE.
3. POSITION ACCURACY OF BLOCKS IN HEAD $\pm .005$
4. POSITION ACCURACY OF SEISMIC MASS WITHIN ACCELEROMETER $\pm .03$ (PER ENDEVCO SPEC.)
5. CENTER OF GRAVITY (C-G) LOCATIONS PER G.M. DRAWINGS 78051-61 AND 78051-338 FOR HYBRID III TEST DUMMY HEAD

D ADD 7264C		23AUG00		SCALE: .75	DRAWN BY: C.R.M./S.E.F.	
C WAS DWG B1846.DWG		13OCT99		DATE: 6-28-85	MODEL: 1846	
B REDRAWN IN CAD		6/30/93		UNLESS OTHERWISE SPECIFIED		
A ADD REAR MOUNT		12/7/89		SURFACE FINISH 125		
SYM		REVISION		TWO PLACE DECIMALS $\pm .010$		
DATE				THREE PLACE DECIMALS $\pm .005$		
				ANGLES $\pm 0^\circ 30'$		
				REMOVE ALL BURRS & SHARP EDGES		
				DIMENSION LIMITS HELD AFTER PLATING		
				ROBERT A. DENTON, INC.		
				ROCHESTER HILLS, MICHIGAN		
				DRAWING NUMBER: B-1846-D		